



NQAD/CAD LINEARITY URACIL

Application note # QAN-0046

PARAMETERS

NQAD

GAS: Air
FILTER: 5 sec
GAIN: 2-500X
EVAPORATION TEMP: off

CAD

GAS: Nitrogen
FILTER: High
RANGE: 2-500 pA

HPLC

MOBILE PHASE: 40:60 water:ACN
FLOW RATE: 1.2 mL/min
COLUMN: Alltech Alltima C18, 3u,
4.6 x 33 mm
COLUMN TEMP: Ambient
INJECTION VOLUME :10 µL
ANALYTE: Uracil



**QUANT TECHNOLOGIES,
LLC**

1463 94TH Lane NE
Blaine, MN, USA 55449
763-398-0508
www.Quant-NQAD.com

LINEARITY DATA

